

GÖLCÜK TERSANESİ KOMUTANLIĞI MEKANİK KALİBRASYON LABORATUVARI İMKÂN VE KABİLİYETLERİ			
KALİBRASYON KONUSU/SAHASI	KULLANILAN KALİBRATÖR MARKA/MODEL	ÖLÇÜM ARALIĞI	HASSASİYET
Akış/Flow	MOLBOX RFM-M	10-600 kPa	± 0.0001
	molbloc-l	0-10 slm	± 0.0001
	molbloc-l	0-100 slm	± 0.0001
	molbloc-s	0-300 slm	± 0.0001
	molbloc-s	0-1200 slm	± 0.0001
	BRONKHORST MASS FLOWMETRE	0-50 l/min	± 0.1 %FS
Basınç	MASTER GAGE (HEISE)	0-15 Psi	± 0.025 Psi
	MASTER GAGE (HEISE)	0-100 Psi	± 0.01 Psi
	MASTER GAGE (HEISE)	0-300 Psi	± 0.5 Psi
	MASTER GAGE (HEISE)	0-1000 Psi	± 1 Psi
	MASTER GAGE (HEISE)	0-3000 Psi	± 5 Psi
	MASTER GAGE (HEISE)	0-10000 Psi	± 25 Psi
	DEAD WEIGHT TESTER (M 2800/3)	10-16000 Psi	0.01 Psi IV
	PRESSURE/VACUM CHAMBER (Model 3558)	50mic absolute/1000 psi	± 0.02psig
	AUTOMATIC PRES.CAL.SYS. (Model 3666)	0-25 inchg vac/10000 psig	± 0.1 % IV
	PRECISION PRES.STANDART (Model 3689A)	0-30 inchg vac/10000 psig	± 0.025 % IV
	DIGITAL BASINÇ GEYİCİ additel 680	0-4200 Bar	± 0.1 Bar
	DIGITAL BASINÇ GEYİCİ DPI 104	0-700 Bar	± 0.1 Bar
	DIGITAL BASINÇ GEYİCİ DPI 104	0-70 Bar	± 0.01 Bar
	DIGITAL BASINÇ GEYİCİ DPI 104	0-7 Bar	± 0.001 Bar
PORTABLE GAGE CALIBRATOR (TYPE 1127-2)	0-30 Inchg 100 PSI	± 0.001 inchg	
Boyutsal	GAGE BLOCK COMPARATOR (Model 136B-3V)	0.3 inch - 12 inch	5 µ inch
	UNIVERSAL MEASURING MACHINE MODEL U 304395	0 inch - 12 inch	30 µ inch
		12 inch - 20 inch	50 µ inch
		20 inch - 40 inch	70 µ inch
		ANGLE BLOCK SET	0-99 °
	MAHR 826 (Gauge Block Comparator)	0-175mm	0.01µ
	GAGE BLOCK SET SS8A1X	0-20 "	GRADE 2
	GAGE BLOCK SET SS88A1X	0-4 "	GRADE 2
GAGE BLOCK SET 516-937-60	0-100mm	GRADE K	
Devir	TACHOMETER CALIBRATION (Model N-11 ECS/3A)	0 – 40000 RPM	± 0.02 % RPM
Gaz	GAZ DEDEKTÖRLERİ KALİBRASYON MODÜLÜ (DRAGER X-DOCK 6300)	GAZ DEDEKTÖRLERİ	
	STANDART KALİBRASYON GAZLARI	C12 5 ppm,	± 1 % SERTİFİKA DEĞERİ
		CH4 2,5 % Vol	± 1 % SERTİFİKA DEĞERİ
		O2 19%	± 1 % SERTİFİKA DEĞERİ
		H2 1 %-1,9 %	± 1 % SERTİFİKA DEĞERİ
		CO2 2 %, 50 %	± 1 % SERTİFİKA DEĞERİ
		CO 100 ppm, 250 ppm	± 1 % SERTİFİKA DEĞERİ
H2S 75 ppm	± 1 % SERTİFİKA DEĞERİ		
Kuvvet	MACHINE, 100000 Lb.Universal CA MODEL A99870-1	0-1000 Lb	± 0.05 % fs
		0-3000 Lb	± 0.05 % fs
		0-10000 Lb	± 0.05 % fs
		0-25000 Lb	± 0.05 % fs
		0-100000 Lb	± 0.05 % fs
Kütle	WEIGHT SET (7228-3)	1 mg - 100 gr	0.01mg E2
	WEIGHT SET (508B)	1 - 5000 gr	0.02mg F1
	WEIGHT SET (7210-3)	1 - 5000 gr	0.02mg E2
	WEIGHT SET (01Lb-50Lb)	0.01 - 50 Libre	± 0.1mg E2
	UVE AĞIRLIK SETİ(1mg-20Kg)	1mg-20Kg	± 0.1mg E2
	MODEL R50	0-2500 Lb	±2 gr
	HCE 25	0-25 kg	±2,5 mgr
	MC210S	0-210gr	± 0.01 mg
	C10000	0-10000 gr	±0.001 gr
	RADWAG 30	0-30 Kg	±0,1 mg
	M1 SINIFI	20 KG. / 2.000 KG.	± 100 mg.
Sıcaklık	MODEL 3604	100 °F / 1199 °F	± 1 °F
	MODEL 3605	-40 °F / 250 °F	± 0.1 °F
	FLUKE 6022	40 °C / 300 °C	± 0,005 °C
	FLUKE 7381	-80 °C / 110 °C	± 0,006 °C
	FLUKE 9102S	-10 °C / 122 °C	± 0,25 °C
	MODEL 1100 CF	-454 °F / 2502 °F	± 5 Microvolt (DC)
	FLUKE 4181 IR KALİBRATÖR	35°C/500°C	±0.35 °C
	DEW POINT MONITOR (Higro M3)	20 - 100 °C	± 0,15 °C IV
		0 - 40 % Rh	± 0,5 % Rh
		41 - 100 % Rh	± 1 % Rh
	FLUKE 9150	150 °C / 1200 °C	± 0.05 °C
	NUVE ID 301	-40°C/150°C	±0.1°C
	FLUKE 3125 (YÜZEY TERMOMETRE KALİB.)	18-98 %Rh	±1 Rh
	FLUKE SPRT MODÜLÜ	35-400 °C	± 1°C
SPRT 162CE	0-400 Ω	0.00001Ω	
	-40°C/650°C	±0.005 °C	
Tork	TORK FORCE CALIBRATOR MODEL 3695	0-50 Lbft	± 0,5 Lbft %
		0-500 Lbft	± 0,5 Lbft %
		0-5000 Lbft	± 0,5 Lbft %
		0-20000 Lbft	± 0,5 Lbft %